

Molecular Innovations, Inc.

46430 Peary Court
Novi, MI 48377
TEL: (248) 896-0142
FAX: (248) 896-0148
<http://www.mol-innov.com>

Certificate of Analysis

Product: Human Plasminogen Total Antigen ELISA Kit – HPLGKT-TOT

Lot: HPLGKT-TOT-315

Date of Manufacture: 23 March 2015

Expiration Date: 31 March 2016

Components:

Instruction/Reagent Sheet	Number of copies: 1	Lot #: HPLGKT-TOT-315
ELISA Plate	Number of plates: 1	Lot #: 913
10X Wash Buffer	Number of vials: 1	Lot #: 215A
1000ng/ml Standard	Number of vials: 1	Lot #: 713L
Primary Antibody	Number of vials: 1	Lot #: 713L
Secondary Antibody	Number of vials: 1	Lot #: DAS-HRP-315
Substrate	Number of vials: 1	Lot #: 14079-A

Test

Average Background OD
Average High OD
Development Time



Specification

≤ 0.15
 ≥ 1.5
 ≤ 10 minutes

Result

0.044
3.71
2 minutes

Quality Control Standard Curve:



Duane E. Day

Duane E. Day