

## Molecular Innovations, Inc.

46430 Peary Court  
Novi, MI 48377  
TEL: (248) 896-0142  
FAX: (248) 896-0148  
<http://www.mol-innov.com>

# Certificate of Analysis

**Product:** Active mouse PAI-1 functional assay ELISA kit – MPAIKT

**Lot:** MPAIKT-415

**Date of Manufacture:** 28 April 2015

**Expiration Date:** 31 October 2016

### Components:

Instruction/Reagent Sheet	Number of copies: 1	Lot #: MPAIKT-415
ELISA Plate	Number of plates: 1	Lot #: 415C
10X Wash Buffer	Number of vials: 1	Lot #: 415A
250ng/ml Standard	Number of vials: 1	Lot #: 415L
Primary Antibody	Number of vials: 1	Lot #: 415L
Secondary Antibody	Number of vials: 1	Lot #: GAR-HRP-1214
Substrate	Number of vials: 1	Lot #: 14349-A

### Test

Average Background OD  
Average High OD  
Development Time



### Specification

$\leq 0.15$   
 $\geq 1.5$   
 $\leq 10$  minutes

### Result

0.071  
4.11  
3.5 minutes

### Quality Control Standard Curve:



*Duane E. Day*

Duane E. Day