

## Molecular Innovations, Inc.

46430 Peary Court  
Novi, MI 48377  
TEL: (248) 896-0142  
FAX: (248) 896-0148  
<http://www.mol-innov.com>

# Certificate of Analysis

**Product:** Mouse tPA total antigen assay ELISA kit – MTPAKT-TOT

**Lot:** MTPAKT-TOT-1013

**Date of Manufacture:** 15 October 2013

**Expiration Date:** 31 October 2015

### Components:

Instruction/Reagent Sheet	Number of copies: 1	Lot #: MTPAKT-TOT-1013
ELISA Plate	Number of plates: 1	Lot #: 812
10X Wash Buffer	Number of vials: 1	Lot #: 1013
1000ng/ml Standard	Number of vials: 1	Lot #: 811L
Primary Antibody	Number of vials: 1	Lot #: 512L
Secondary Antibody	Number of vials: 1	Lot #: GAM-HRP-713
Substrate	Number of vials: 1	Lot #: 13233-A

### Test

Average Background OD  
Average High OD  
Development Time



### Specification

$\leq 0.15$   
 $\geq 1.5$   
 $\leq 10$  minutes

### Result

0.068  
3.79  
2.5 minutes

### Quality Control Standard Curve:



*Duane E. Day*

Duane E. Day