

## Molecular Innovations, Inc.

46430 Peary Court  
Novi, MI 48377  
TEL: (248) 896-0142  
FAX: (248) 896-0148  
<http://www.mol-innov.com>

# Certificate of Analysis

**Product:** Mouse uPA total antigen assay ELISA kit – MUPAKT-TOT

**Lot:** MUPAKT-TOT-316

**Date of Manufacture:** 16 March 2016



**Expiration Date:** 31 March 2017

### Components:

Instruction/Reagent Sheet	Number of copies: 1	Lot #: MUPAKT-TOT-316
ELISA Plate	Number of plates: 1	Lot #: 316C
10X Wash Buffer	Number of vials: 1	Lot #: 316A
100ng/ml Standard	Number of vials: 1	Lot #: 314L
Primary Antibody	Number of vials: 1	Lot #: 414L
Secondary Antibody	Number of vials: 1	Lot #: GAR-HRP-316
Substrate	Number of vials: 1	Lot #: 15299-A

Test	Specification	Result
Average Background OD	$\leq 0.15$	0.084
Average High OD	$\geq 1.5$	2.85
Development Time	$\leq 10$ minutes	4 minutes

### Quality Control Standard Curve:



*Duane E. Day*

Duane E. Day