

Molecular Innovations, Inc.

46430 Peary Court
Novi, MI 48377
TEL: (248) 896-0142
FAX: (248) 896-0148
<http://www.mol-innov.com>

Certificate of Analysis

Product: Active rat PAI-1 functional assay ELISA kit – RPAIKT

Lot: RPAIKT-415

Date of Manufacture: 16 April 2015

Expiration Date: 28 April 2016

Components:

Instruction/Reagent Sheet	Number of copies: 1	Lot #: RPAIKT-415
ELISA Plate	Number of plates: 1	Lot #: REO-415C
10X Wash Buffer	Number of vials: 1	Lot #: 415
250ng/ml Standard	Number of vials: 1	Lot #: 315L
Primary Antibody	Number of vials: 1	Lot #: 414L
Secondary Antibody	Number of vials: 1	Lot #: GAR-HRP-1214
Substrate	Number of vials: 1	Lot #: 14349-A

Test

Average Background OD
Average High OD
Development Time



Specification

≤ 0.15
 ≥ 1.5
 ≤ 10 minutes

Result

0.066
4.23
3 minutes

Quality Control Standard Curve:



Duane E. Day

Duane E. Day