

Molecular Innovations, Inc.

46430 Peary Court
Novi, MI 48377
TEL: (248) 896-0142
FAX: (248) 896-0148
<http://www.mol-innov.com>

Certificate of Analysis

Product: SARS-CoV-2 IgG Seroconversion ELISA Kit – SARSCOV2GKT **Lot:** 520-C

Date of Manufacture: 20 May 2020

Expiration Date: 30 May 2021

Components:

Instruction/Reagent Sheet	Number of copies: 1	Lot #: 520-C
ELISA Plate	Number of plates: 1	Lot #: 520C-A
10X Wash Buffer	Number of vials: 1	Lot #: 420
Sample Buffer	Number of vials: 1	Lot #: 520
Negative Control	Number of vials: 1	Lot #: 520L-A
Positive Control	Number of vials: 1	Lot #: 520L-C
Secondary Antibody	Number of vials: 1	Lot #: 520L
Substrate	Number of vials: 1	Lot #: 19355-A
Stop Solution	Number of vials: 1	Lot #: 167926

Test	Specification	Result
Average Background OD	≤ 0.15	0.042
Average Negative Control	≤ 0.3	0.154
Average Negative Control	≥ 0.5	0.772

Duane E. Day

Duane E. Day